



CERTIFICATE OF DECONTAMINATION

LuxarCare LLC, 16932 Woodinville Redmond Rd NE A109, Woodinville, WA 98072

Toll Free: 866-589-2722 * Ph: 425-487-9988 * Fax: 425-487-1157

NOTE: OSHA Standards (29CFR1910.1030) Require Equipment be Decontaminated Prior to Shipping

OSHA Bloodborne Pathogens Standard (29CFR1910.1030) requires the use of an appropriate chemical disinfectant on potentially contaminated equipment prior to transporting. This is necessary to ensure that equipment is in a condition that makes it safe to be handled by all personnel who may come into contact with it during transit and subsequent handling. All used surgical equipment is assumed to have potential exposure to blood or other potentially infectious materials (OPIM).

Ensure the following before packaging for shipment

- All returned equipment must be decontaminated or sterilized.
- Lasers, fibers, and non-sterilizable accessories must be properly decontaminated for shipping.
- Used accessories such as handpieces must be sterilized for shipping
- Do NOT return any used disposables including single use tips, smoke evacuator filters or hoses
- Special LuxarCare permission and instructions are required IF:
 - it is suspected that blood or OPIM has contaminated the INSIDE of any equipment.
 - used accessories, including disposables, are being returned in response to a product failure, complaint, warranty claim, or where decontamination or sterilization may influence a potential investigation.

Instructions

1. Decontaminate or sterilize using OSHA-approved protocols. Decontamination solutions typically require ten minutes processing time before wiping down. There must be no visible blood or OPIM present when decontamination is complete.
2. Complete below checklist to ensure typical problem areas have been addressed.
3. Carefully and correctly package. For proper Packaging and Material Instructions you can visit our website or speak with a Customer Support Representative for assistance. Failure to do so could result in a delay in processing. The customer is responsible for any shipping damage that results from improper packaging.

- () Decontamination of equipment for shipment to LuxarCare was performed as above
- () Special Approval by LuxarCare Received by _____ (LC Customer Service Representative Initials)

CASE # _____

(Case numbers may be obtained from LuxarCare Customer Service at 866-589-2722)

NOTE: Failure to follow OSHA decontamination standards or failure to include a signed and dated decontamination form (ML103) will result in a \$125.00 fee.

Laser Tower; Laser Footswitch and Cord; Top of Laser Base; Underneath Laser Base; Wheels.

Decontamination for shipment to LuxarCare LLC was performed by:

Printed Name: _____ Signature: _____

Facility Name: _____ Phone: _____ Date: _____

ML103 - Decontamination Form